

# MIF

- MIF was evaluated using Supplier B's Human Sepsis/Apoptosis
- For 30 serum samples from healthy donors, 10% had undetectable MIF, while the remainder ranged from 21.2 to 626.2 pg/ml

# Suppliers

- A: Bio-Rad
- B: Linco Research (done)
- C: Upstate
- D: Biosource
- E: RnD Systems